

ANKAphase: software for single-distance phase retrieval from inline X-ray phase-contrast radiographs. Erratum


T. Weitkamp, D. Haas, D. Wegrzynek and A. Rack

J. Synchrotron Rad. (2013). **20**, 205

Copyright © International Union of Crystallography

Author(s) of this paper may load this reprint on their own web site or institutional repository provided that this cover page is retained. Reproduction of this article or its storage in electronic databases other than as specified above is not permitted without prior permission in writing from the IUCr.

For further information see <http://journals.iucr.org/services/authorrights.html>



Synchrotron radiation research is rapidly expanding with many new sources of radiation being created globally. Synchrotron radiation plays a leading role in pure science and in emerging technologies. The *Journal of Synchrotron Radiation* provides comprehensive coverage of the entire field of synchrotron radiation research including instrumentation, theory, computing and scientific applications in areas such as biology, nanoscience and materials science. Rapid publication ensures an up-to-date information resource for scientists and engineers in the field.

Crystallography Journals **Online** is available from journals.iucr.org

***ANKA*phase: software for single-distance
 phase retrieval from inline X-ray phase-
 contrast radiographs. Erratum**

**T. Weitkamp,^{a*} D. Haas,^{b,c} D. Wegrzynek^{d,‡} and
 A. Rack^{e*}**

^aSynchrotron Soleil, Gif-sur-Yvette, France, ^bInstitute for Synchrotron Radiation/
 ANKA Light Source, KIT, Karlsruhe, Germany, ^cHochschule Darmstadt, Darmstadt,
 Germany, ^dInternational Atomic Energy Agency, Seibersdorf, Austria, and ^eEuropean
 Synchrotron Radiation Facility, Grenoble, France.
 E-mail: weitkamp@synchrotron-soleil.fr, arack@snaflu.de

An equation in the paper by Weitkamp *et al.* [(2011),
J. Synchrotron Rad. **18**, 617–629] is corrected.

There was a mistake in equation (6) in the paper by Weitkamp *et al.*
 (2011). The equation should read

$$\varphi(x, y) = \frac{\delta}{2\beta} \ln \left(\mathcal{F}^{-1} \left\{ \frac{\mathcal{F}[I(x, y)/I_0(x, y)]}{1 + [\lambda z \delta / (4\pi\beta)](u^2 + v^2)} \right\} \right). \quad (6)$$

References

Weitkamp, T., Haas, D., Wegrzynek, D. & Rack, A. (2011). *J. Synchrotron Rad.*
18, 617–629.

[‡] Present address: University of Science and Technology, Kraków, Poland.